

**Central Glass & Ceramic Research Institute
KOLKATA (WEST BENGAL) INDIA
CORRIGENDUM**

REFERENCE NO. :- P/NC/21/SS/SO(SP)/OT/21-22

DATE: 04/08/2021

NAME OF EQUIPMENT: GLASS CUTTING MACHINE

NOTE: The Bids must be submitted in the Central Public Procurement Portal (URL:<https://etenders.gov.in/eprocure/app>) only. Manual/Offline bids shall not be accepted under any circumstances. Bidders should quote in INR only.

CONSEQUENT TO THE PRE-BID MEETING HELD ON 28/07/2021, THE REVISED SPECIFICATION IS GIVEN BELOW:-

SPECIFICATIONS:-

Final Specification : Cutting machine after Pre-bid meeting

| Sl. No | Details of Items* | Quantity |
|--------|---|-----------|
| 1 | <p>Glass cutting machine (A precision section saw designed for cutting all types of glasses with fixture and table)</p> <p>1. End Use: Low-Speed Precision cutter is applicable to accurate cutting of hard materials, especially to high-value fragile glass.</p> <p>2. Detailed specification:</p> <ol style="list-style-type: none"> I. Cutting capacity up to 20 mm – 4" wheel II. Cutting capacity up to 30 mm – 5" wheel III. Positional accuracy: 0.01 mm IV. Arbor size: 12.7 mm V. Aluminium based and stainless steel construction door VI. The machine should have limit switch. VII. Automatic Cut-Off switch provision VIII. Micro-meter for cross-feed adjustment (0 -25 mm) IX. Movable Re-Circulation coolant tray (Made of SS) X. Coolant Splash Guard (Transparent) facility XI. Single Phase DC Motor (within 100 Watts) XII. Colour touch screen HMI XIII. Down-feed facility with variable weights: 400 gm (Maximum) XIV. Variable speed : 60-600 rpm XV. Emergency Stop button XVI. Spare/additional diamond wheel (4" diameter and 0.3 mm thick)- 1 no XVII. Spare/ additional diamond wheel (5" diameter and 0.4 mm thick) – 1no XVIII. Synthetic Lubricating Oil – 5 liter XIX. Suitable granite top table with sink for placement of cutting machine <p>3. After sales, services to be provided by the supplier during warranty period on one year.</p> <p>4. Inspection and Tests required: No.</p> <p>5. Acceptance test: Smooth functioning of equipment.</p> | 01 No. |

The above amendments shall amount to amendments of the relevant terms of our Bid Document for CGCRI Tender No. **P/NC/21/SS/SO(SP)/OT/21-22** .

All other Tender terms and conditions remain unchanged.

Bidders should quote only in INR.

Sd-
(R. Ray)
Controller of Stores & Purchase

*****#