



सीएसआईआर - केन्द्रीय काँच एवं सिरामिक अनुसंधान संस्थान
196, राजा एस सी मल्लिक रोड, कोलकाता - 700 032, भारत
CSIR - CENTRAL GLASS & CERAMIC RESEARCH INSTITUTE
196, Raja S C Mullick Road, Kolkata - 700 032, India



CORRIGENDUM

REFERENCE NO. : - P/NC/69/NCP/SO(SB)/OTE/22-23

DATE: 31/08/2022

NAME OF EQUIPMENT: "SUPPLY, INSTALLATION, COMMISSIONING, DEMONSTRATION & TRAINING OF ATMOSPHERE CONTROLLED 1100 ° C CHAMBER FURNACE"

NOTE: The Bids must be submitted in the Central Public Procurement Portal (URL:<https://etenders.gov.in/eprocure/app>) only. Manual/Offline bids shall not be accepted under any circumstances. Bidders should quote in INR only.

CONSEQUENT TO THE PRE-BID MEETING HELD ON 29/08/2022, THE REVISED SPECIFICATION IS GIVEN BELOW:-

SPECIFICATIONS:-

- 1) Rotameters: 2 (two) nos. of precision Rotameters (Gas flow- controllers, make Eureka or equivalent) should be provided and they must be calibrated for – (i) one for N₂, (ii) one for H₂ (5%)-Argon mixtures.
- 2) Warranty: 24 months form date of successful commissioning of furnace. Warranty shall also cover the replacement of fault/damage of heating elements if occurred within the warranty tenure.
- 3) Layout/drawing of the Sample tray is attached in Annexure-'A'

Deadline for Submission of Bid and Opening of Bid is remain unchanged as given in CGCRI Tender No. P/NC/69/NCP/SO(SB)/OTE/22-23 dated 18/08/2022 which is as follows:-

Bid Submission End Date & Time	17/09/2022 upto 3.30 AM(IST)
Bid Opening Date & Time	19/09/2022 upto 3.30 PM(IST)

The above amendments shall amount to amendments of the relevant terms of our Bid Document for CGCRI Tender No. P/NC/69/NCP/SO(SB)/OTE/22-23 dated 18/08/2022.

All other Tendered technical specifications and terms & conditions will be remain unchanged.

Bidders should quote only in INR. Bidding is open to only Class I /Class II Local Bidders.


अंजनी कुमार पाण्डेय / Anjani Kumar Pandey
भंडारण एवं क्रय अधिकारी / Stores & Purchase Officer
सीएसआईआर - केन्द्रीय काँच एवं सिरामिक अनुसंधान संस्थान
CSIR - CENTRAL GLASS & CERAMIC RESEARCH INSTITUTE
196, राजा एस सी मल्लिक रोड / 196, Raja S. C. Mullick Road
कोलकाता / Kolkata- 700 032

31/08/2022
(A.K.Pandey)
Stores & Purchase Officer
FOR & ON BEHALF OF CSIR

#*****#

Annexure-A:

'Layout/drawing for Inconel sample Tray and top/cover lid



31.08.2012

अंजनी कुमार पाण्डेय / Anjani Kumar Pandey
बुखार एवं क्रय अधिकारी / Stores & Purchase Officer
सीएसआईआर - केन्द्रीय काँच एवं सिरेमिक अनुसंधान संस्थान
CSIR - CENTRAL GLASS & CERAMIC RESEARCH INSTITUTE
198, राजा एस. सी. मुल्लिक रोड / 198, Raja S. C. Mullick Road
कोलकाता / Kolkata- 700 032